
Midmark M11 - v1.0.5
Total Cycles: 4148

Name: RITTER2
Sterilizer ID: V2313228
11/21/2024 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.7 F 0.0
2:00 170.8 F 2.4
4:00 202.1 F 6.4
6:00 199.8 F 10.7
8:00 218.8 F 5.1
10:00 229.1 F 8.5
12:00 242.4 F 13.3
14:00 239.8 F 11.1
16:00 237.7 F 9.9
18:00 251.0 F 16.2
20:00 262.1 F 22.5
22:00 272.1 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.6
2:00 271.6 F 28.8
3:00 271.8 F 29.5
4:00 273.0 F 29.9

Min 271.0 F 28.7
Max 273.1 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.7 F 0.0
5:00 214.7 F 0.0
10:00 191.8 F 0.0
15:00 187.4 F 0.0
20:00 182.9 F 0.0
25:00 179.0 F 0.0
30:00 175.9 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:52
DRYING: 30:00
TOTAL CYCLE: 01:01:48
CYCLE COMPLETE

11/21/2024 01:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____