
Midmark M11 - v1.0.5
Total Cycles: 2893

Name: RITTER2
Sterilizer ID: V2313228
09/15/2023 04:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 158.7 F 0.0
2:00 155.5 F 2.4
4:00 177.1 F 6.0
6:00 196.6 F 9.8
8:00 209.8 F 10.7
10:00 227.5 F 7.4
12:00 236.0 F 11.7
14:00 233.6 F 8.5
16:00 241.4 F 12.1
18:00 244.4 F 12.7
20:00 255.4 F 18.6
22:00 265.5 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.8
1:00 272.9 F 29.8
2:00 272.7 F 29.9
3:00 271.9 F 29.7
4:00 271.6 F 29.2

Min 270.9 F 28.7
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.3 F 0.0
5:00 227.1 F 0.0
10:00 225.2 F 0.0
15:00 184.9 F 0.0
20:00 174.9 F 0.0
25:00 173.5 F 0.0
30:00 173.4 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 23:50
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:02:52
CYCLE COMPLETE

09/15/2023 05:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____