
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 005471
15:03 02/05/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:41
PRESSURIZING 1:45
STERILIZING 2:44
136.4 C 223kPa 2:44
136.4 C 223kPa 3:14
136.1 C 223kPa 3:44
136.4 C 224kPa 4:14
136.4 C 224kPa 4:44
136.3 C 222kPa 5:14
136.4 C 225kPa 5:44
136.4 C 225kPa 6:14
136.4 C 225kPa 6:14
VENTING 6:14
AIR DRYING 7:06
CYCLE COMPLETE 7:15

Digital Signature #
7CCB881BB2015B6C
