
Midmark M11 - v1.0.0
Total Cycles: 3132

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 03:03 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 186.2 F 0.0
2:00 184.4 F 0.0
4:00 201.4 F 5.1
6:00 218.1 F 10.4
8:00 226.1 F 6.3
10:00 238.6 F 11.8
12:00 237.8 F 9.7
14:00 238.3 F 8.8
16:00 250.8 F 15.5
18:00 263.0 F 22.4
20:00 271.9 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.1
1:00 272.6 F 28.8
2:00 271.4 F 28.0
3:00 272.4 F 28.9
4:00 272.3 F 28.5

Min 271.2 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.0 F 0.0
5:00 210.9 F 0.0
10:00 222.3 F 0.0
15:00 223.5 F 0.0
20:00 223.1 F 0.0
25:00 221.2 F 0.0
30:00 216.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 20:10
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 00:58:31
CYCLE COMPLETE

04/19/2023 04:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____