
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001474
11:59 25/09/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:12
PRESSURIZING 3:14
STERILIZING 4:21
136.4 C 224kPa 4:21
136.3 C 225kPa 4:51
136.4 C 227kPa 5:21
136.4 C 227kPa 5:51
136.3 C 225kPa 6:21
135.9 C 224kPa 6:51
136.3 C 225kPa 7:21
136.4 C 226kPa 7:51
136.4 C 226kPa 7:51
VENTING 7:52
AIR DRYING 8:46
CYCLE COMPLETE 10:15

Digital Signature #
17D3E27E239ABA74
