
Midmark M11 - v1.0.5
Total Cycles: 135

Name: RITTER2
Sterilizer ID: V2313228
06/23/2021 03:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.8 F 0.0
2:00 172.5 F 3.5
4:00 192.4 F 8.8
6:00 216.2 F 7.3
8:00 227.3 F 10.6
10:00 232.7 F 8.7
12:00 238.6 F 9.8
14:00 248.9 F 15.1
16:00 261.7 F 22.4
18:00 272.2 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 271.9 F 29.6
2:00 272.0 F 29.3
3:00 271.8 F 29.1
4:00 271.6 F 29.0

Min 271.3 F 28.9
Max 272.3 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.4 F 0.0
5:00 228.9 F 0.0
10:00 222.7 F 0.0
15:00 231.0 F 0.0
20:00 222.5 F 0.0
25:00 229.2 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:38
HEATING: 18:08
STERILIZING: 4:00
VENTING: 2:52
DRYING: 30:00
TOTAL CYCLE: 00:56:38
CYCLE COMPLETE

06/23/2021 04:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____