

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1986

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 11:03 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.7 F 0.0  
2:00 130.8 F 1.8  
4:00 156.2 F 4.6  
6:00 193.7 F 8.8  
8:00 201.8 F 13.5  
10:00 217.5 F 6.1  
12:00 228.1 F 10.6  
14:00 231.1 F 7.0  
16:00 240.6 F 11.9  
18:00 238.0 F 9.6  
20:00 250.7 F 15.8  
22:00 261.8 F 22.2  
24:00 272.4 F 29.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.5  
1:00 272.5 F 29.2  
2:00 271.5 F 28.5  
3:00 272.1 F 29.2  
4:00 272.2 F 28.9

Min 271.1 F 28.3  
Max 272.6 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 207.7 F 0.0  
5:00 216.2 F 0.0  
10:00 226.9 F 0.0  
15:00 236.4 F 0.0  
20:00 227.1 F 0.0  
25:00 231.5 F 0.0  
30:00 223.7 F 0.0  
Drying Complete

FILLING: 0:51  
HEATING: 24:30  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:02:32  
CYCLE COMPLETE

06/01/2022 12:05 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_