

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1831

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2022 11:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 138.4 F 0.0  
2:00 141.2 F 0.0  
4:00 153.3 F 0.2  
6:00 178.0 F 5.0  
8:00 199.1 F 9.7  
10:00 213.6 F 7.6  
12:00 227.5 F 9.7  
14:00 236.1 F 9.2  
16:00 242.6 F 13.3  
18:00 243.4 F 12.5  
20:00 256.3 F 19.4  
22:00 267.7 F 26.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 29.2  
1:00 272.9 F 29.8  
2:00 271.9 F 29.2  
3:00 273.0 F 30.0  
4:00 272.4 F 29.4

Min 271.7 F 29.1  
Max 273.1 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.0 F 0.0  
5:00 187.2 F 0.0  
10:00 199.1 F 0.0  
15:00 206.2 F 0.0  
20:00 210.4 F 0.0  
25:00 204.2 F 0.0  
30:00 204.6 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 23:20  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:36  
CYCLE COMPLETE

12/13/2022 12:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_