
Midmark M11 - v1.0.5
Total Cycles: 2924

Name: RITTER2
Sterilizer ID: V2313228
09/25/2023 06:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.4 F 0.0
2:00 175.8 F 3.1
4:00 194.9 F 7.7
6:00 212.3 F 12.3
8:00 224.3 F 7.0
10:00 233.1 F 12.0
12:00 236.5 F 9.4
14:00 238.7 F 10.0
16:00 248.1 F 14.4
18:00 259.6 F 20.7
20:00 269.9 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.6 F 29.1
2:00 271.2 F 28.4
3:00 272.4 F 29.3
4:00 271.9 F 28.8

Min 271.0 F 28.4
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.1 F 0.0
5:00 218.7 F 0.0
10:00 232.0 F 0.0
15:00 227.7 F 0.0
20:00 224.7 F 0.0
25:00 223.7 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 2:36
HEATING: 20:58
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:01:15
CYCLE COMPLETE

09/25/2023 07:03 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____