
Midmark M11 - v1.0.0
Total Cycles: 850

Name: RITTER 1
Sterilizer ID: V2178198
05/31/2021 09:28 AM

BEGIN PACKS CYCLE
Temp: 250 Degrees F
Time: 30 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 154.5 F 0.0
2:00 154.7 F 1.6
4:00 165.6 F 4.5
6:00 196.3 F 2.9
8:00 211.4 F 3.8
10:00 224.9 F 5.6
12:00 228.4 F 6.1
14:00 227.4 F 5.1
16:00 241.5 F 10.8
18:00 252.7 F 16.2

STERILIZING
mm:ss Degrees PSI
0:00 251.3 F 15.7
1:00 252.5 F 16.1
2:00 251.3 F 15.8
3:00 252.7 F 16.2
4:00 251.6 F 15.7
5:00 252.0 F 16.2
6:00 252.6 F 16.2
7:00 252.0 F 15.9
8:00 251.3 F 15.6
9:00 251.9 F 16.2
10:00 252.7 F 16.3
11:00 252.4 F 16.1
12:00 252.2 F 15.9
13:00 251.6 F 15.7
14:00 251.2 F 15.7
15:00 251.8 F 16.1
16:00 252.5 F 16.3
17:00 252.6 F 16.3
18:00 252.5 F 16.2
19:00 252.3 F 16.1
20:00 252.2 F 15.9
21:00 251.7 F 15.8
22:00 251.5 F 15.7
23:00 251.3 F 15.7
24:00 251.6 F 16.0
25:00 252.2 F 16.3
26:00 252.7 F 16.4
27:00 252.6 F 16.3
28:00 252.5 F 16.2
29:00 252.4 F 16.1
30:00 252.1 F 16.0

Min 251.2 F 15.6
Max 252.8 F 16.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.2 F 0.0
5:00 190.9 F 0.0
10:00 210.5 F 0.0
15:00 214.4 F 0.0
20:00 219.9 F 0.0
25:00 210.9 F 0.0
30:00 182.4 F 0.0
Drying Complete

FILLING: 1:02

HEATING: 18:19
STERILIZING: 30:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:22:18
CYCLE COMPLETE

05/31/2021 10:50 AM
ACCEPTED: X
REJECTED:
Operator: U01