

-----  
Midmark M11 - v1.0.5  
Total Cycles: 793

Name: RITTER2  
Sterilizer ID: V2313228  
02/14/2022 10:48 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 160.1 F 0.0  
2:00 156.5 F 2.7  
4:00 171.9 F 6.8  
6:00 199.9 F 12.7  
8:00 228.4 F 7.9  
10:00 236.8 F 11.6  
12:00 240.8 F 11.8  
14:00 240.3 F 10.8  
16:00 253.6 F 17.9  
18:00 266.2 F 25.2  
20:00 272.8 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.4  
1:00 272.8 F 29.5  
2:00 272.7 F 29.9  
3:00 271.6 F 29.5  
4:00 272.8 F 29.6  
  
Min 271.4 F 29.2  
Max 273.2 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 180.0 F 0.0  
5:00 226.2 F 0.0  
10:00 215.5 F 0.0  
15:00 195.4 F 0.0  
20:00 218.0 F 0.0  
25:00 217.2 F 0.0  
30:00 232.0 F 0.0  
Drying Complete

FILLING: 1:17  
HEATING: 20:19  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 00:58:41  
CYCLE COMPLETE

02/14/2022 11:46 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_