

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1560

Name: RITTER2  
Sterilizer ID: V2313228  
10/04/2022 11:05 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 204.6 F 0.0  
2:00 197.9 F 0.3  
4:00 211.9 F 4.8  
6:00 229.9 F 10.3  
8:00 233.2 F 7.3  
10:00 242.4 F 13.4  
12:00 243.0 F 12.2  
14:00 242.9 F 11.8  
16:00 256.5 F 19.0  
18:00 268.3 F 26.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.6  
1:00 272.5 F 29.3  
2:00 271.4 F 28.5  
3:00 272.5 F 29.4  
4:00 272.0 F 28.9  
  
Min 271.3 F 28.5  
Max 272.6 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.8 F 0.0  
5:00 223.7 F 0.0  
10:00 241.4 F 0.0  
15:00 227.4 F 0.0  
20:00 223.8 F 0.0  
25:00 222.4 F 0.0  
30:00 219.7 F 0.0  
Drying Complete

FILLING: 1:36  
HEATING: 19:14  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 00:57:59  
CYCLE COMPLETE

10/04/2022 12:02 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_