
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008264
9:26 29/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:57
PRESSURIZING 2:05
STERILIZING 3:00
136.0 C 220kPa 3:00
135.9 C 219kPa 3:30
135.8 C 218kPa 4:00
136.3 C 219kPa 4:30
136.4 C 222kPa 5:00
136.2 C 222kPa 5:30
136.2 C 221kPa 6:00
136.4 C 219kPa 6:30
136.4 C 219kPa 6:30
VENTING 6:30
AIR DRYING 7:19
CYCLE COMPLETE 27:14

Digital Signature #
88A44054B86B8858
