
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 007677
7:35 16/07/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:52
PRESSURIZING 3:52
STERILIZING 4:56
136.0 C 220kPa 4:56
136.4 C 220kPa 5:26
136.3 C 223kPa 5:56
136.4 C 222kPa 6:26
136.1 C 221kPa 6:56
136.4 C 223kPa 7:26
136.3 C 221kPa 7:56
136.2 C 221kPa 8:26
136.2 C 221kPa 8:26
VENTING 8:26
AIR DRYING 9:17
CYCLE COMPLETE 20:57

Digital Signature #
1E1E9491AC493D88
