
Midmark M11 - v1.0.5
Total Cycles: 25

Name: RITTER2
Sterilizer ID: V2313228
04/27/2021 10:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.8 F 0.0
2:00 169.1 F 3.4
4:00 186.8 F 7.6
6:00 207.6 F 12.6
8:00 219.4 F 7.2
10:00 232.6 F 12.5
12:00 234.8 F 10.7
14:00 237.4 F 9.8
16:00 251.0 F 16.7
18:00 262.7 F 23.6
20:00 272.9 F 30.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.6
1:00 272.7 F 30.2
2:00 271.7 F 29.5
3:00 272.5 F 30.4
4:00 272.4 F 29.9

Min 271.6 F 29.4
Max 272.8 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.3 F 0.0
5:00 232.1 F 0.0
10:00 231.9 F 0.0
15:00 224.0 F 0.0
20:00 221.0 F 0.0
25:00 219.4 F 0.0
30:00 217.4 F 0.0
Drying Complete

FILLING: 1:49
HEATING: 20:17
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:59:19
CYCLE COMPLETE

04/27/2021 11:50 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____