
Midmark M11 - v1.0.5
Total Cycles: 223

Name: RITTER2
Sterilizer ID: V2313228
07/19/2021 11:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.8 F 0.0
2:00 183.2 F 3.1
4:00 198.3 F 8.1
6:00 215.9 F 13.2
8:00 228.7 F 8.9
10:00 239.8 F 11.1
12:00 241.5 F 12.5
14:00 241.9 F 11.8
16:00 254.5 F 18.4
18:00 265.5 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 272.7 F 29.7
2:00 271.7 F 29.2
3:00 272.8 F 29.9
4:00 271.8 F 29.1

Min 271.7 F 29.0
Max 273.0 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.1 F 0.0
5:00 211.1 F 0.0
10:00 214.0 F 0.0
15:00 216.6 F 0.0
20:00 194.6 F 0.0
25:00 187.6 F 0.0
30:00 220.2 F 0.0

Drying Complete

FILLING: 2:23
HEATING: 19:47
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:59:27
CYCLE COMPLETE

07/19/2021 12:26 PM
ACCEPTED: X
REJECTED:
Operator: U01