

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1165

Name: RITTER2  
Sterilizer ID: V2313228  
06/07/2022 09:06 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.3 F 0.0  
2:00 190.6 F 3.3  
4:00 200.0 F 8.8  
6:00 229.7 F 8.5  
8:00 235.8 F 10.8  
10:00 233.9 F 8.6  
12:00 240.1 F 10.3  
14:00 247.4 F 14.3  
16:00 259.2 F 21.0  
18:00 270.2 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.5  
1:00 272.1 F 29.6  
2:00 271.9 F 29.7  
3:00 272.0 F 29.6  
4:00 272.2 F 29.4  
  
Min 271.8 F 29.3  
Max 272.9 F 30.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.7 F 0.0  
5:00 232.4 F 0.0  
10:00 228.5 F 0.0  
15:00 231.6 F 0.0  
20:00 224.5 F 0.0  
25:00 223.1 F 0.0  
30:00 224.7 F 0.0  
Drying Complete

FILLING: 2:28  
HEATING: 19:27  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:05  
CYCLE COMPLETE

06/07/2022 10:08 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02