
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000991
9:24 14/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:37
PRESSURIZING 2:42
STERILIZING 3:50
136.4 C 233kPa 3:50
136.4 C 233kPa 4:20
136.4 C 230kPa 4:50
136.4 C 231kPa 5:20
136.4 C 233kPa 5:50
136.3 C 233kPa 6:20
136.2 C 231kPa 6:50
136.4 C 233kPa 7:20
136.4 C 233kPa 7:20
VENTING 7:20
AIR DRYING 8:12
CYCLE COMPLETE 8:31

Digital Signature #
E4248CFB8D4B5188
