

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4234

Name: RITTER 1  
Sterilizer ID: V2178198  
01/26/2024 09:20 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.8 F 0.0  
2:00 70.5 F 0.2  
4:00 82.7 F 1.2  
6:00 114.8 F 3.4  
8:00 146.9 F 6.2  
10:00 173.1 F 9.5  
12:00 192.4 F 13.1  
14:00 206.2 F 3.6  
16:00 219.6 F 6.7  
18:00 226.3 F 10.0  
20:00 234.5 F 10.7  
22:00 234.6 F 9.2  
24:00 242.1 F 11.2  
26:00 243.0 F 11.8  
28:00 254.7 F 17.6  
30:00 264.7 F 23.6  
32:00 271.1 F 27.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.4  
1:00 272.7 F 28.7  
2:00 272.3 F 28.9  
3:00 272.2 F 28.4  
4:00 272.8 F 29.1  
  
Min 271.5 F 28.3  
Max 272.9 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 174.4 F 0.0  
5:00 240.0 F 0.0  
10:00 229.9 F 0.0  
15:00 218.3 F 0.0  
20:00 217.5 F 0.0  
25:00 221.8 F 0.0  
30:00 226.8 F 0.0  
  
Drying Complete

FILLING: 0:52  
HEATING: 32:47  
STERILIZING: 4:00  
VENTING: 3:25  
DRYING: 30:00  
TOTAL CYCLE: 01:11:04  
CYCLE COMPLETE

01/26/2024 10:31 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_