
Midmark M11 - v1.0.0
Total Cycles: 1879

Name: RITTER 1
Sterilizer ID: V2178198
04/27/2022 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 203.6 F 0.0
2:00 200.1 F 3.6
4:00 205.9 F 8.2
6:00 217.3 F 8.7
8:00 230.4 F 8.8
10:00 241.0 F 10.9
12:00 241.0 F 11.6
14:00 239.3 F 10.1
16:00 253.0 F 16.9
18:00 264.8 F 23.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.7 F 29.0
2:00 271.5 F 28.3
3:00 272.7 F 29.1
4:00 271.8 F 28.4

Min 271.5 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.2 F 0.0
5:00 212.9 F 0.0
10:00 223.5 F 0.0
15:00 228.1 F 0.0
20:00 230.7 F 0.0
25:00 227.2 F 0.0
30:00 224.3 F 0.0
Drying Complete

FILLING: 1:26
HEATING: 19:51
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:26
CYCLE COMPLETE

04/27/2022 11:38 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____