
Midmark M11 - v1.0.0
Total Cycles: 3126

Name: RITTER 1
Sterilizer ID: V2178198
04/19/2023 08:27 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 184.6 F 0.0
4:00 202.1 F 3.5
6:00 220.8 F 9.0
8:00 229.5 F 9.0
10:00 237.1 F 10.8
12:00 236.7 F 8.5
14:00 240.3 F 9.9
16:00 248.4 F 14.3
18:00 260.8 F 21.2
20:00 272.2 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.5
2:00 271.9 F 28.6
3:00 272.1 F 28.4
4:00 271.8 F 28.6

Min 271.4 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 221.8 F 0.0
10:00 237.0 F 0.0
15:00 227.2 F 0.0
20:00 225.9 F 0.0
25:00 227.2 F 0.0
30:00 227.3 F 0.0
Drying Complete

FILLING: 1:45
HEATING: 20:28
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:31
CYCLE COMPLETE

04/19/2023 09:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____