
Midmark M11 - v1.0.0
Total Cycles: 2324

Name: RITTER 1
Sterilizer ID: V2178198
09/14/2022 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 177.6 F 4.2
4:00 189.6 F 8.6
6:00 206.5 F 13.5
8:00 223.0 F 8.0
10:00 235.6 F 13.2
12:00 237.2 F 10.1
14:00 234.9 F 8.1
16:00 248.5 F 14.9
18:00 260.9 F 21.5
20:00 271.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.4 F 29.1
2:00 272.6 F 29.5
3:00 272.2 F 29.4
4:00 272.2 F 29.0

Min 271.9 F 28.9
Max 272.6 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.4 F 0.0
5:00 216.0 F 0.0
10:00 227.7 F 0.0
15:00 231.0 F 0.0
20:00 229.8 F 0.0
25:00 227.0 F 0.0
30:00 227.8 F 0.0
Drying Complete

FILLING: 1:04
HEATING: 21:22
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:59:35
CYCLE COMPLETE

09/14/2022 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____