
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000768
11:40 28/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:35
PRESSURIZING 2:40
STERILIZING 3:48
136.0 C 228kPa 3:48
136.3 C 233kPa 4:18
136.4 C 232kPa 4:48
136.1 C 231kPa 5:18
136.1 C 230kPa 5:48
136.4 C 233kPa 6:18
136.3 C 231kPa 6:48
136.4 C 233kPa 7:18
136.4 C 233kPa 7:18
VENTING 7:18
AIR DRYING 8:13
CYCLE COMPLETE 8:15

Digital Signature #
71A095A3AB42F15D
