
Midmark M11 - v1.0.5
Total Cycles: 563

Name: RITTER2
Sterilizer ID: V2313228
11/18/2021 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 106.4 F 0.0
2:00 107.1 F 0.0
4:00 133.3 F 0.1
6:00 167.4 F 2.7
8:00 184.8 F 7.0
10:00 203.1 F 11.6
12:00 217.9 F 5.9
14:00 229.1 F 10.8
16:00 231.0 F 7.6
18:00 243.2 F 12.7
20:00 243.6 F 12.6
22:00 255.7 F 19.1
24:00 266.9 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.2
1:00 272.4 F 29.6
2:00 271.6 F 29.6
3:00 272.1 F 29.4
4:00 271.9 F 29.8

Min 271.3 F 29.1
Max 272.5 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 183.0 F 0.0
5:00 202.5 F 0.0
10:00 219.3 F 0.0
15:00 226.1 F 0.0
20:00 227.4 F 0.0
25:00 216.5 F 0.0
30:00 208.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 25:27
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 01:03:45
CYCLE COMPLETE

11/18/2021 10:46 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____