
Midmark M11 - v1.0.0
Total Cycles: 5012

Name: RITTER 1
Sterilizer ID: V2178198
10/21/2024 02:34 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 77.4 F 0.0
2:00 88.8 F 0.0
4:00 126.8 F 0.0
6:00 126.0 F 0.1
8:00 157.3 F 0.4
10:00 184.2 F 1.7
12:00 206.6 F 4.9
14:00 214.0 F 8.2
16:00 224.0 F 11.9
18:00 225.3 F 6.4
20:00 237.2 F 11.1
22:00 233.0 F 7.6
24:00 245.6 F 12.6
26:00 244.1 F 12.1
28:00 256.0 F 18.3
30:00 266.6 F 24.7

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.0
1:00 272.1 F 28.7
2:00 272.1 F 28.3
3:00 271.7 F 28.3
4:00 272.6 F 28.7

Min 271.4 F 27.9
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 192.3 F 0.0
10:00 194.9 F 0.0
15:00 208.1 F 0.0
20:00 222.7 F 0.0
25:00 235.5 F 0.0
30:00 234.6 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 31:29
STERILIZING: 4:00
VENTING: 3:30
DRYING: 30:00
TOTAL CYCLE: 01:09:43
CYCLE COMPLETE

10/21/2024 03:43 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____