
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.5uS / 0.3ppm
CYCLE NUMBER 000502
12:07 07/05/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:11
PRESSURIZING 3:15
STERILIZING 4:24
136.0 C 230kPa 4:24
136.3 C 232kPa 4:54
136.4 C 233kPa 5:24
136.2 C 232kPa 5:54
136.3 C 233kPa 6:24
136.3 C 233kPa 6:54
136.4 C 233kPa 7:24
136.1 C 231kPa 7:54
136.1 C 231kPa 7:54
VENTING 7:54
AIR DRYING 8:48
CYCLE COMPLETE 9:28

Digital Signature #
370E402B2391B0D4
