

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000722  
8:03 22/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:27  
STERILIZING 3:30  
136.0 C 228kPa 3:30  
136.1 C 230kPa 4:00  
136.4 C 230kPa 4:30  
136.1 C 229kPa 5:00  
136.3 C 231kPa 5:30  
136.2 C 230kPa 6:00  
136.2 C 231kPa 6:30  
136.4 C 232kPa 7:00  
136.4 C 232kPa 7:00  
VENTING 7:00  
AIR DRYING 7:51  
CYCLE COMPLETE 7:58

-----  
Digital Signature #  
47663CFCA43E4EEC  
-----