
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.9uS / 0.5ppm
CYCLE NUMBER 001129
8:46 21/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:33
PRESSURIZING 4:33
STERILIZING 5:58
136.4 C 223kPa 5:58
136.3 C 223kPa 6:28
136.4 C 224kPa 6:58
136.2 C 222kPa 7:28
136.4 C 224kPa 7:58
136.3 C 224kPa 8:28
136.4 C 224kPa 8:58
136.2 C 223kPa 9:28
136.2 C 223kPa 9:28
VENTING 9:28
AIR DRYING 10:22
CYCLE COMPLETE 25:22

Digital Signature #
6C1588C3CF729782
