
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000380
9:02 01/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:43
PRESSURIZING 3:45
STERILIZING 4:59
136.5 C 224kPa 4:59
136.2 C 222kPa 5:29
136.4 C 224kPa 5:59
136.1 C 221kPa 6:29
136.2 C 222kPa 6:59
136.4 C 221kPa 7:29
136.4 C 224kPa 7:59
136.4 C 225kPa 8:29
136.4 C 225kPa 8:29
VENTING 8:30
AIR DRYING 9:22
CYCLE COMPLETE 10:34

Digital Signature #
DD384FB3D1BB2C4A
