
Midmark M11 - v1.0.5
Total Cycles: 1351

Name: RITTER2
Sterilizer ID: V2313228
07/26/2022 03:24 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 166.4 F 0.0
2:00 168.9 F 0.2
4:00 186.8 F 3.4
6:00 208.3 F 8.6
8:00 221.2 F 9.8
10:00 232.1 F 10.3
12:00 235.7 F 8.4
14:00 241.7 F 11.2
16:00 246.9 F 14.2
18:00 259.5 F 21.2
20:00 270.8 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.8 F 29.8
2:00 271.8 F 29.0
3:00 272.8 F 30.0
4:00 272.4 F 29.4

Min 271.6 F 29.0
Max 272.9 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 200.0 F 0.0
10:00 211.9 F 0.0
15:00 221.4 F 0.0
20:00 221.2 F 0.0
25:00 214.4 F 0.0
Dry Cycle Aborted

FILLING: 1:49
HEATING: 20:47
STERILIZING: 4:00
VENTING: 3:11
DRYING: 27:43
TOTAL CYCLE: 00:57:30

07/26/2022 04:21 PM
ACCEPTED: X
REJECTED:
Operator: U02