
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 002588
13:25 21/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:04
PRESSURIZING 2:09
STERILIZING 3:07
136.4 C 226kPa 3:07
136.3 C 224kPa 3:37
136.2 C 226kPa 4:07
136.3 C 222kPa 4:37
136.1 C 223kPa 5:07
136.3 C 224kPa 5:37
136.3 C 224kPa 6:07
136.3 C 222kPa 6:37
136.3 C 222kPa 6:37
VENTING 6:38
AIR DRYING 7:29
CYCLE COMPLETE 37:29

Digital Signature #
2F66EFDAC4EA317B
