
Midmark M11 - v1.0.0
Total Cycles: 4680

Name: RITTER 1
Sterilizer ID: V2178198
06/26/2024 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.1 F 0.0
2:00 153.1 F 2.8
4:00 169.3 F 6.0
6:00 198.9 F 10.2
8:00 222.3 F 6.9
10:00 227.4 F 7.8
12:00 237.0 F 12.2
14:00 233.2 F 8.5
16:00 243.4 F 12.1
18:00 243.2 F 12.2
20:00 255.0 F 18.1
22:00 265.5 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 272.5 F 29.4
2:00 272.0 F 29.0
3:00 272.6 F 29.6
4:00 272.1 F 29.1

Min 270.9 F 28.4
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.1 F 0.0
5:00 219.1 F 0.0
10:00 245.9 F 0.0
15:00 233.9 F 0.0
20:00 215.1 F 0.0
25:00 214.5 F 0.0
30:00 203.2 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:51
DRYING: 30:00
TOTAL CYCLE: 01:02:45
CYCLE COMPLETE

06/26/2024 03:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____