
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 008347
10:15 25/02/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:31
STERILIZING 3:32
136.0 C 220kPa 3:32
136.2 C 223kPa 4:02
136.4 C 220kPa 4:32
136.4 C 221kPa 5:02
136.4 C 222kPa 5:32
136.3 C 220kPa 6:02
136.4 C 224kPa 6:32
136.4 C 222kPa 7:02
136.4 C 222kPa 7:02
VENTING 7:03
AIR DRYING 7:58
CYCLE COMPLETE 8:11

Digital Signature #
22BB1C8B56A8D8F1
