

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1376

Name: RITTER 1  
Sterilizer ID: V2178198  
11/17/2021 11:50 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.3 F 0.0  
2:00 181.6 F 3.0  
4:00 210.9 F 7.2  
6:00 223.8 F 12.8  
8:00 229.4 F 7.9  
10:00 242.7 F 13.2  
12:00 240.5 F 10.6  
14:00 236.7 F 8.7  
16:00 251.1 F 15.6  
18:00 262.8 F 22.4  
20:00 272.4 F 28.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.4  
1:00 272.1 F 28.4  
2:00 272.8 F 28.7  
3:00 272.6 F 29.0  
4:00 271.6 F 28.3

Min 271.5 F 28.3  
Max 272.8 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.7 F 0.0  
5:00 225.1 F 0.0  
10:00 209.7 F 0.0  
15:00 208.6 F 0.0  
20:00 211.7 F 0.0  
25:00 223.1 F 0.0  
30:00 222.8 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 20:54  
STERILIZING: 4:00  
VENTING: 3:09  
DRYING: 30:00  
TOTAL CYCLE: 00:59:35  
CYCLE COMPLETE

11/17/2021 12:49 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_