
Midmark M11 - v1.0.0
Total Cycles: 2134

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2022 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 182.0 F 0.0
2:00 178.3 F 0.7
4:00 179.5 F 4.3
6:00 211.3 F 8.5
8:00 223.2 F 8.9
10:00 232.9 F 8.8
12:00 236.5 F 10.8
14:00 238.8 F 11.0
16:00 236.5 F 8.8
18:00 250.0 F 15.4
20:00 261.8 F 21.9
22:00 272.1 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.4 F 29.2
2:00 271.5 F 28.3
3:00 271.9 F 29.1
4:00 272.2 F 28.8

Min 271.2 F 28.2
Max 272.7 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.0 F 0.0
5:00 238.8 F 0.0
10:00 234.3 F 0.0
15:00 227.1 F 0.0
20:00 218.5 F 0.0
25:00 225.2 F 0.0
30:00 224.6 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 22:32
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:00:35
CYCLE COMPLETE

07/15/2022 09:19 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____