
Midmark M11 - v1.0.0
Total Cycles: 5095

Name: RITTER 1
Sterilizer ID: V2178198
11/14/2024 07:07 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 170.5 F 0.0
2:00 170.9 F 0.4
4:00 192.1 F 1.0
6:00 208.3 F 4.6
8:00 224.6 F 8.6
10:00 235.8 F 13.0
12:00 233.7 F 8.9
14:00 240.1 F 11.0
16:00 243.6 F 12.3
18:00 242.0 F 11.1
20:00 254.5 F 17.4
22:00 265.3 F 23.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.0
1:00 272.4 F 28.7
2:00 271.5 F 27.9
3:00 272.4 F 28.8
4:00 271.8 F 28.3

Min 271.3 F 27.9
Max 272.6 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.7 F 0.0
5:00 200.7 F 0.0
10:00 198.7 F 0.0
15:00 202.1 F 0.0
20:00 225.3 F 0.0
25:00 229.7 F 0.0
30:00 230.9 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 23:53
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:02:58
CYCLE COMPLETE

11/14/2024 08:09 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____