
Midmark M11 - v1.0.0
Total Cycles: 3638

Name: RITTER 1
Sterilizer ID: V2178198
08/11/2023 10:00 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.3 F 0.0
2:00 73.0 F 0.0
4:00 105.0 F 0.0
6:00 129.2 F 0.0
8:00 155.0 F 0.3
10:00 185.9 F 3.7
12:00 204.0 F 7.7
14:00 224.9 F 11.8
16:00 221.5 F 5.5
18:00 231.9 F 9.4
20:00 238.9 F 11.1
22:00 241.0 F 11.1
24:00 237.3 F 9.1
26:00 251.8 F 16.0
28:00 263.6 F 22.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.2
1:00 272.0 F 28.5
2:00 272.5 F 28.9
3:00 272.8 F 28.9
4:00 272.8 F 28.8

Min 271.6 F 28.0
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 235.5 F 0.0
Dry Cycle Aborted

FILLING: 0:57
HEATING: 29:56
STERILIZING: 4:00
VENTING: 2:50
DRYING: 0:16
TOTAL CYCLE: 00:37:60

08/11/2023 10:37 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____