

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2027

Name: RITTER 1  
Sterilizer ID: V2178198  
06/16/2022 08:35 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.4 F 0.0  
2:00 140.3 F 0.7  
4:00 148.3 F 2.1  
6:00 173.5 F 5.4  
8:00 195.4 F 10.0  
10:00 223.9 F 5.1  
12:00 230.2 F 8.4  
14:00 240.0 F 13.2  
16:00 239.3 F 10.3  
18:00 235.4 F 7.9  
20:00 248.4 F 14.3  
22:00 259.7 F 20.5  
24:00 269.7 F 27.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 271.5 F 28.6  
2:00 271.7 F 28.6  
3:00 271.9 F 28.6  
4:00 272.2 F 28.8

Min 271.3 F 28.5  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.0 F 0.0  
5:00 226.3 F 0.0  
10:00 227.0 F 0.0  
15:00 225.5 F 0.0  
20:00 235.4 F 0.0  
25:00 225.7 F 0.0  
30:00 223.4 F 0.0  
Drying Complete

FILLING: 0:49  
HEATING: 25:36  
STERILIZING: 4:00  
VENTING: 3:05  
DRYING: 30:00  
TOTAL CYCLE: 01:03:30  
CYCLE COMPLETE

06/16/2022 09:38 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_