
Midmark M11 - v1.0.5
Total Cycles: 3418

Name: RITTER2
Sterilizer ID: V2313228
02/28/2024 12:08 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.2 F 0.0
2:00 105.3 F 0.0
4:00 112.5 F 0.6
6:00 136.3 F 3.0
8:00 164.1 F 6.5
10:00 196.4 F 10.1
12:00 221.4 F 8.2
14:00 220.2 F 7.9
16:00 229.6 F 12.3
18:00 231.9 F 9.0
20:00 239.1 F 11.3
22:00 243.2 F 12.9
24:00 254.6 F 18.8
26:00 265.2 F 25.1
28:00 270.7 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 30.1
1:00 272.9 F 30.5
2:00 271.7 F 30.2
3:00 272.8 F 30.4
4:00 271.6 F 30.2

Min 271.5 F 29.9
Max 273.0 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.7 F 0.0
5:00 212.1 F 0.0
10:00 230.0 F 0.0
15:00 225.3 F 0.0
20:00 211.8 F 0.0
25:00 208.4 F 0.0
30:00 207.0 F 0.0
Drying Complete

FILLING: 0:53
HEATING: 28:48
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:07:19
CYCLE COMPLETE

02/28/2024 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____