
Midmark M11 - v1.0.0
Total Cycles: 3622

Name: RITTER 1
Sterilizer ID: V2178198
08/08/2023 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 172.1 F 0.0
2:00 170.8 F 0.0
4:00 172.9 F 1.7
6:00 190.6 F 5.6
8:00 206.8 F 9.5
10:00 219.8 F 10.5
12:00 225.9 F 7.7
14:00 235.9 F 12.0
16:00 234.7 F 8.8
18:00 242.7 F 11.7
20:00 244.1 F 12.4
22:00 255.1 F 18.1
24:00 265.1 F 24.0
26:00 271.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.3
1:00 272.6 F 29.1
2:00 271.9 F 28.4
3:00 271.6 F 28.6
4:00 272.7 F 29.0

Min 271.4 F 28.2
Max 272.8 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.0 F 0.0
5:00 212.5 F 0.0
10:00 216.9 F 0.0
15:00 217.2 F 0.0
20:00 216.0 F 0.0
25:00 213.7 F 0.0
30:00 213.6 F 0.0
Drying Complete

FILLING: 0:55
HEATING: 26:03
STERILIZING: 4:00
VENTING: 3:34
DRYING: 30:00
TOTAL CYCLE: 01:04:32
CYCLE COMPLETE

08/08/2023 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____