
Midmark M11 - v1.0.5
Total Cycles: 2526

Name: RITTER2
Sterilizer ID: V2313228
06/16/2023 09:14 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 177.4 F 0.0
2:00 173.7 F 2.3
4:00 179.1 F 6.8
6:00 203.4 F 12.6
8:00 224.5 F 8.1
10:00 236.9 F 9.6
12:00 243.8 F 12.5
14:00 245.8 F 13.5
16:00 260.0 F 21.2
18:00 258.2 F 20.1
20:00 272.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.3 F 28.8
2:00 273.0 F 29.3
3:00 272.4 F 29.3
4:00 272.1 F 28.7

Min 271.8 F 28.5
Max 273.0 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 224.6 F 0.0
Dry Cycle Aborted

FILLING: 1:32
HEATING: 20:19
STERILIZING: 4:00
VENTING: 2:57
DRYING: 1:42
TOTAL CYCLE: 00:30:29

06/16/2023 09:44 AM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____