

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1987

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2022 12:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 142.5 F 0.0  
2:00 144.4 F 2.7  
4:00 168.6 F 6.1  
6:00 186.1 F 10.4  
8:00 211.6 F 4.6  
10:00 219.9 F 8.6  
12:00 232.4 F 13.3  
14:00 234.3 F 9.5  
16:00 237.3 F 9.3  
18:00 245.3 F 13.1  
20:00 257.0 F 19.4  
22:00 267.8 F 26.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.7 F 29.5  
2:00 272.1 F 28.9  
3:00 271.6 F 29.0  
4:00 272.7 F 29.5

Min 271.4 F 28.6  
Max 272.8 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 202.4 F 0.0  
5:00 214.0 F 0.0  
10:00 230.5 F 0.0  
15:00 230.0 F 0.0  
20:00 225.2 F 0.0  
25:00 228.0 F 0.0  
30:00 225.5 F 0.0  
Drying Complete

FILLING: 0:55  
HEATING: 23:25  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 01:01:31  
CYCLE COMPLETE

06/01/2022 01:42 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_