

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 007511  
9:13 07/06/2024

UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:39  
PRESSURIZING 2:41  
STERILIZING 3:50  
136.5 C 220kPa 3:50  
136.0 C 217kPa 4:20  
136.4 C 222kPa 4:50  
136.4 C 220kPa 5:20  
136.3 C 223kPa 5:50  
136.4 C 222kPa 6:20  
136.4 C 223kPa 6:50  
136.4 C 222kPa 7:20  
136.4 C 222kPa 7:20  
VENTING 7:20  
AIR DRYING 8:09  
CYCLE COMPLETE 38:09

-----

Digital Signature #  
97A93DBA2800B536  
-----