
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 006457
7:47 13/10/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:04
PRESSURIZING 3:07
STERILIZING 4:14
136.4 C 222kPa 4:14
136.0 C 220kPa 4:44
135.9 C 220kPa 5:14
136.1 C 221kPa 5:44
136.2 C 220kPa 6:14
136.2 C 221kPa 6:44
136.0 C 219kPa 7:14
136.2 C 221kPa 7:44
136.2 C 221kPa 7:44
VENTING 7:44
AIR DRYING 8:33
CYCLE COMPLETE 38:33

Digital Signature #
2E8DC3FE43DF5142
