

-----  
Midmark M11 - v1.0.0  
Total Cycles: 636

Name: MIDMARK  
Sterilizer ID: V2178198  
01/27/2021 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 117.5 F 0.0  
2:00 118.5 F 0.4  
4:00 134.7 F 2.3  
6:00 169.2 F 5.9  
8:00 199.6 F 10.9  
10:00 216.6 F 5.1  
12:00 226.1 F 11.1  
14:00 229.6 F 7.2  
16:00 241.2 F 13.3  
18:00 241.6 F 11.3  
20:00 254.7 F 18.0  
22:00 266.4 F 25.0  
24:00 272.1 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 28.8  
1:00 271.9 F 29.3  
2:00 272.5 F 29.2  
3:00 272.3 F 29.0  
4:00 271.9 F 28.8

Min 271.7 F 28.7  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 204.2 F 0.0  
5:00 237.6 F 0.0  
10:00 229.3 F 0.0  
15:00 231.7 F 0.0  
20:00 220.1 F 0.0  
25:00 212.7 F 0.0  
30:00 208.7 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 24:04  
STERILIZING: 4:00  
VENTING: 2:60  
DRYING: 30:00  
TOTAL CYCLE: 01:02:05  
CYCLE COMPLETE

01/27/2021 02:33 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_