
Midmark M11 - v1.0.5
Total Cycles: 1901

Name: RITTER2
Sterilizer ID: V2313228
01/05/2023 01:44 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 165.6 F 0.0
2:00 163.2 F 3.3
4:00 178.1 F 7.0
6:00 200.3 F 11.8
8:00 216.0 F 5.9
10:00 232.1 F 10.9
12:00 231.4 F 7.9
14:00 243.0 F 12.3
16:00 243.4 F 12.6
18:00 255.4 F 19.0
20:00 266.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 29.4
1:00 272.5 F 30.0
2:00 271.6 F 29.3
3:00 272.1 F 29.8
4:00 271.6 F 29.4

Min 271.5 F 29.2
Max 272.5 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.0 F 0.0
5:00 232.0 F 0.0
10:00 230.3 F 0.0
15:00 227.6 F 0.0
20:00 230.8 F 0.0
25:00 231.7 F 0.0
30:00 227.1 F 0.0
Drying Complete

FILLING: 1:21
HEATING: 21:31
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:59:57
CYCLE COMPLETE

01/05/2023 02:43 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____