
Midmark M11 - v1.0.5
Total Cycles: 3955

Name: RITTER2
Sterilizer ID: V2313228
09/18/2024 05:25 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.3 F 0.0
2:00 139.5 F 1.7
4:00 152.4 F 4.7
6:00 173.9 F 8.6
8:00 195.9 F 13.2
10:00 226.6 F 7.3
12:00 233.0 F 12.2
14:00 233.6 F 9.1
16:00 239.5 F 10.4
18:00 246.4 F 13.7
20:00 257.5 F 20.1
22:00 268.2 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.5
1:00 271.3 F 29.4
2:00 271.8 F 29.5
3:00 273.0 F 30.9
4:00 272.4 F 30.2

Min 270.9 F 29.4
Max 274.2 F 30.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.8 F 0.0
5:00 231.5 F 0.0
10:00 217.6 F 0.0
15:00 184.9 F 0.0
20:00 174.0 F 0.0
25:00 171.7 F 0.0
30:00 173.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 23:55
STERILIZING: 4:00
VENTING: 3:46
DRYING: 30:00
TOTAL CYCLE: 01:02:53
CYCLE COMPLETE

09/18/2024 06:27 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____