
Midmark M11 - v1.0.0
Total Cycles: 4005

Name: RITTER 1
Sterilizer ID: V2178198
11/10/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.9 F 0.0
2:00 129.6 F 1.1
4:00 136.4 F 3.5
6:00 164.6 F 7.1
8:00 196.6 F 11.5
10:00 210.2 F 4.4
12:00 223.3 F 8.7
14:00 233.0 F 13.3
16:00 234.9 F 9.5
18:00 240.3 F 9.7
20:00 245.3 F 13.1
22:00 257.0 F 19.2
24:00 267.6 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.6 F 28.4
3:00 271.6 F 28.7
4:00 272.6 F 29.1

Min 271.1 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 208.6 F 0.0
5:00 212.9 F 0.0
10:00 231.8 F 0.0
15:00 230.8 F 0.0
20:00 229.2 F 0.0
25:00 226.2 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 25:25
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:03:48
CYCLE COMPLETE

11/10/2023 11:23 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____