
Midmark M11 - v1.0.0
Total Cycles: 5326

Name: RITTER 1
Sterilizer ID: V2178198
02/10/2025 10:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.3 F 0.0
2:00 72.1 F 0.4
4:00 101.8 F 2.1
6:00 136.2 F 4.5
8:00 150.2 F 7.2
10:00 170.8 F 10.6
12:00 184.7 F 5.9
14:00 204.1 F 4.2
16:00 216.3 F 7.5
18:00 227.5 F 11.3
20:00 226.2 F 5.9
22:00 239.3 F 11.3
24:00 235.9 F 8.5
26:00 249.4 F 14.9
28:00 260.8 F 21.3
30:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 272.1 F 28.3
1:00 272.3 F 28.8
2:00 272.6 F 28.7
3:00 271.8 F 28.4
4:00 272.9 F 29.0

Min 271.6 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 201.6 F 0.0
5:00 188.6 F 0.0
10:00 197.7 F 0.0
15:00 205.1 F 0.0
20:00 209.2 F 0.0
25:00 210.3 F 0.0
30:00 213.5 F 0.0
Drying Complete

FILLING: 0:43
HEATING: 30:33
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:49
CYCLE COMPLETE

02/10/2025 11:12 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____