
Midmark M11 - v1.0.5
Total Cycles: 1726

Name: RITTER2
Sterilizer ID: V2313228
11/15/2022 12:05 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 188.9 F 0.0
2:00 191.8 F 3.3
4:00 204.1 F 9.1
6:00 220.3 F 7.7
8:00 234.9 F 11.3
10:00 235.8 F 9.5
12:00 238.3 F 9.1
14:00 251.0 F 16.1
16:00 263.4 F 23.4
18:00 271.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.8
1:00 272.5 F 29.3
2:00 271.7 F 29.1
3:00 272.4 F 29.3
4:00 271.6 F 28.8

Min 271.5 F 28.7
Max 272.7 F 29.6

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.7 F 0.0
5:00 206.5 F 0.0
10:00 221.8 F 0.0
15:00 231.2 F 0.0
20:00 230.7 F 0.0
25:00 225.4 F 0.0
30:00 223.9 F 0.0
Drying Complete

FILLING: 2:46
HEATING: 18:00
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:57
CYCLE COMPLETE

11/15/2022 01:06 PM
ACCEPTED: X
REJECTED:
Operator: U02